



**DICOM
CONFORMANCE
STATEMENT
FOR ZIOBASE 4.4**

0.1	APPLICATION DATA FLOW	4
0.2	FUNCTIONAL DEFINITION OF AE'S	5
0.3	SEQUENCING OF REAL-WORLD ACTIVITIES	5
1	AE SPECIFICATIONS	6
1.1	ASSOCIATION POLICIES	7
1.1.1	<i>General</i>	7
1.1.2	<i>Number of Associations</i>	7
1.1.3	<i>Asynchronous Nature</i>	7
1.1.4	<i>Implementation Identifying Information</i>	7
1.2	ASSOCIATION INITIATION POLICY	7
1.2.1	<i>Real-World Activities</i>	7
1.2.1.1	Description and Sequencing of Activities.....	7
1.2.1.2	Proposed Presentation Contexts	8
1.2.1.3	SOP Specific Conformance	13
1.2.1.3.1	SOP Specific Conformance for Verification SOP Class.....	13
1.2.1.3.2	SOP Specific Conformance for Storage SOP Class	13
1.2.1.3.3	SOP Specific Conformance for Query/Retrieve Information Model - FIND	13
1.2.1.3.4	SOP Specific Conformance for Query/Retrieve Information Model - MOVE.....	14
1.2.1.3.5	SOP Specific Conformance for Print Management Service Class	14
1.2.1.3.5.1	Basic Film Session SOP Class.....	14
1.2.1.3.5.2	Basic Film Box SOP Class.....	15
1.2.1.3.5.3	Basic Grayscale Image Box SOP Class	15
1.2.1.3.5.4	Basic Color Image Box SOP Class	16
1.2.1.3.5.5	Printer SOP Class.....	17
1.3	ASSOCIATION ACCEPTANCE POLICY.....	17
1.3.1	<i>Real-World Activities</i>	17
1.3.1.1	Description and Sequencing of Real-World Activities.....	17
1.3.1.2	Proposed Presentation Contexts	17
1.3.1.3	SOP Specific Conformance	22
1.3.1.3.1	SOP Specific Conformance for Verification SOP Class.....	22
1.3.1.3.2	SOP Specific Conformance for Storage SOP Class	22
1.3.1.3.3	SOP Specific Conformance for Query/Retrieve Information Model – FIND.....	23
1.3.1.3.4	SOP Specific Conformance for Query/Retrieve Information Model – MOVE.....	24
1.3.1.4	Proposed Presentation Contexts	24
1.3.1.5	Transfer Syntax Selection Policy.....	24
2	NETWORK PROFILE	24
2.1	SUPPORTED PROTOCOL STACKS (PS 3.8, PS 3.9).....	24
2.2	OSI STACK.....	24
2.3	TCP/IP STACK	24
2.4	API.....	24
2.5	PHYSICAL DEVICE SUPPORT	25
2.6	POINT-TO-POINT STACK	25
3	STANDARD EXTENDED/SPECIALIZED/PRIVATE SOP CLASSES	25
4	CONFIGURATION.....	25
4.1	AE TITLE/PRESENTATION ADDRESS MAPPING	25
4.2	PARAMETERS	25
5	SUPPORT OF CHARACTER SETS.....	25



DICOM CONFORMANCE STATEMENT OVERVIEW

This document is the DICOM Conformance Statement for ZIOBASE 4.4.

ZIOBASE supports the following DICOM services:

- Verification Service Class User (SCU)
- Verification Service Class Provider (SCP)
- Storage Service Class User (SCU)
- Storage Service Class Provider (SCP)
- Query/Retrieve Service Class User (SCU)
- Query/Retrieve Service Class Provider (SCP)
- Print Service Class User (SCU)

- IMPLEMENTATION MODEL

0.1 Application Data Flow




Figure 1: Application Data Flow Diagram



ZIOBASE implementation model is shown in Figure 1.

The remote DICOM Application Entities (AE), which are associated with ZIOBASE, should have been configured by the support engineers beforehand.

The remote DICOM AE can request Image Storage Service (C-STORE), Verification Service (C-ECHO) or Query/Retrieve Service (C-FIND/C-MOVE) to ZIOBASE.

ZIOBASE sends request for the Verification service to the remote AE.

ZIOBASE starts up the association for Query/Retrieve service requests to remote AE.

ZIOBASE sends request for the Print service to the remote AE.

0.2 Functional Definition of AE's

ZIOBASE supports the following functions:

- Receive and accept associations from remote AE
- Initiate associations
- Send DICOM image data (SCU)
- Receive DICOM image data (SCP)
- Send request for Verification service to remote AE (SCU)
- Respond to Verification service from remote AE (SCP)
- Send request for Query/Retrieve service to remote AE (SCU)
- Respond to Query/Retrieve service from remote AE (SCP)
- Send request for Print service to remote AE (SCU)
- Manage image files by using database

0.3 Sequencing of Real-World Activities

Sequencing of Real-World Activities is not applied.



1 AE SPECIFICATIONS

ZIOBASE provides Standard Conformance to the following SOP Classes.

Table 1 : SOP Classes for ZIOBASE

SOP Class Name	SOP Class UID	SCU	SCP
Verification SOP Class	1.2.840.10008.1.1	Yes	Yes
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Yes	Yes
Digital X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Yes	Yes
Digital X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Yes	Yes
Digital Mammography X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Yes	Yes
Digital Mammography X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.2.1	Yes	Yes
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Yes	Yes
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Yes	Yes
Ultrasound Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Yes	Yes
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Yes	Yes
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Yes	Yes
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Yes	Yes
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Yes	Yes
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Yes	Yes
X-Ray Radiofluoroscopic Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Yes	Yes
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Yes	Yes
Positron Emission Tomography Image Storage	1.2.840.10008.5.1.4.1.1.128	Yes	Yes
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Yes	Yes
Comprehensive SR Storage	1.2.840.10008.5.1.4.1.1.88.33	Yes	Yes
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Yes	Yes
Study Root Query/Retrieve Information Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Yes	Yes
Basic Grayscale Print Management Meta SOP Class	1.2.840.10008.5.1.1.9	Yes	No
Basic Color Print Management Meta SOP Class	1.2.840.10008.5.1.1.18	Yes	No



1.1 Association Policies

1.1.1 General

The DICOM standard application context name for DICOM 3.0 is always proposed.

Table 2 : DICOM Application Context

DICOM Application Context Name	1.2.840.10008.3.1.1.1
--------------------------------	-----------------------

Max PDU size is configurable by settings. The default is 65536 (64K).

1.1.2 Number of Associations

The number of simultaneous associations is 64 as default. This value is configurable.

1.1.3 Asynchronous Nature

ZIOBASE does not support asynchronous operations (or sub-operations) window negotiation.

1.1.4 Implementation Identifying Information

The implementation information for this Application Entity is.

Table 3 : DICOM Implementation Class and Version for AE Storage

Implementation Class UID	1.2.392.200080.100.200
Implementation Version Name	ZIO_DCM_SVR_200

1.2 Association Initiation Policy

ZIOBASE initiates the association C-STORE between the remote AE by C-MOVE sub-operation. The remote AE has to be registered on ZIOBASE beforehand.

ZIOBASE also initiates the association between the remote AE at:

- Verification
- Manual Transfer
- Query/Retrieve
- Print

by operator's interaction.

1.2.1 Real-World Activities

Each activity listed on 1.2 initiates the DICOM association.

1.2.1.1 Description and Sequencing of Activities

ZIOBASE initiates the association C-STORE to remote AE which are specified by the C-MOVE service.

At Verification, it sends C-ECHO request to remote AE.

At Manual Transfer, it sends DICOM image data to remote AE.

At Query/Retrieve, it queries about remote AE DICOM image data information, and it receives necessary data.

At Print, it sends the DICOM image data after film's width/height format has been



specified.

1.2.1.2 Proposed Presentation Contexts

ZIOBASE is capable of proposing the Presentation Contexts shown in the following table.

Table 4 : Proposed Presentation Contexts

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
Verification SOP Class	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
Computed Radiography Image Storage	1.2.840.10008.5.1.4.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Digital X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Digital X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical,	1.2.840.10008.1.2.4.70	SCU	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
		First-Order Prediction			
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Digital Mammography X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Digital Mammography X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Ultrasound Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
X-Ray Radiofluoroscopic Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Positron Emission Tomography Image Storage	1.2.840.10008.5.1.4.1.1.128	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
RT Image Storage	1.2.840.10008.5.1.4.1.1.481.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCU	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCU	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCU	None
		RLE Lossless	1.2.840.10008.1.2.5	SCU	None
Comprehensive SR	1.2.840.10008.5.1.4.1.1.88.3	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg
Name	UID	Name List	UID List		
Storage	3	Explicit VR Little Endian	1.2.840.10008.1.2.1	SCU	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCU	None
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
Study Root Query/Retrieve Information Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
Basic Grayscale Print Management Meta SOP Class	1.2.840.100008.5.1.1.9	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None
Basic Color Print Management Meta SOP Class	1.2.840.10008.5.1.1.18	Implicit VR Little Endian	1.2.840.10008.1.2	SCU	None

1.2.1.3 SOP Specific Conformance

1.2.1.3.1 SOP Specific Conformance for Verification SOP Class

ZIOBASE provides standard conformance as SCU of Verification SOP Class.

1.2.1.3.2 SOP Specific Conformance for Storage SOP Class

ZIOBASE can execute multiple C-STORE requests as storage service user with a single association.

1.2.1.3.3 SOP Specific Conformance for Query/Retrieve Information Model - FIND

ZIOBASE does not support relational queries. No extended negotiation is performed. The following table shows available keys for queries.

Table 5 : Study Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type
Study Date	(0008,0020)	R
Study Time	(0008,0030)	R
Accession Number	(0008,0050)	R
Patient's Name	(0010,0010)	R
Patient ID	(0010,0020)	R
Study ID	(0020,0010)	R
Study Instance UID	(0020,000d)	U
Number of Study Related Series	(0020,1206)	O
Number of Study Related Instances	(0020,1208)	O



Attribute Name	Tag	Type
Modalities in Study	(0008,0061)	O
Referring Physician's Name	(0008,0090)	O
Study Description	(0008,1030)	O

Table 6 : Series Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type
Modality	(0008,0060)	R
Series Number	(0020,0011)	R
Series Instance UID	(0020,000e)	U
Study Instance UID	(0020,000d)	U(*)
Number of Series Related Instances	(0020,1209)	O
Series Description	(0008,103e)	O
Body Part Examined	(0018,0015)	O
Protocol Name	(0018,1030)	O

(*): Higher level unique key

Table 7 : Instance Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type
Instance Number	(0020,0013)	R
SOP Instance UID	(0008,0018)	U
Study Instance UID	(0020,000d)	U(*)
Series Instance UID	(0020,000e)	U(*)

(*): Higher level unique key

1.2.1.3.4 SOP Specific Conformance for Query/Retrieve Information Model - MOVE

ZIOBASE does not support relational retrievals. No extended negotiation is performed.

1.2.1.3.5 SOP Specific Conformance for Print Management Service Class

Switching between Color and Grayscale is manual. No extended negotiation is performed.

ZIOBASE supports the following SOP classes as defined by the Print management Service Class.

Table 8 : SOP Classes for Print Management Service Class

SOP Class Name	SOP Class UID
Basic Film Session SOP Class	1.2.840.10008.5.1.1.1
Basic Film Box SOP Class	1.2.840.10008.5.1.1.2
Basic Grayscale Image Box SOP Class	1.2.840.10008.5.1.1.4
Basic Color Image Box SOP Class	1.2.840.10008.5.1.1.4.1
Printer SOP Class	1.2.840.10008.5.1.1.16

1.2.1.3.5.1 Basic Film Session SOP Class

The ZIOBASE provides the following support for the Basic Film Session attributes sent by the N-CREATE DIMSE service's SCU.



Table 9 : Basic Film Session SOP Class Request Attributes

Attribute	Tag	Valid Range
Number of Copies	(2000,0010)	1 - 1000
Print Priority	(2000,0020)	MED
Medium Type	(2000,0030)	CLEAR FILM BLUE FILM PAPER
Film Destination	(2000,0040)	MAGAZINE PROCESSOR BIN_1 BIN_2 BIN_3 BIN_4 BIN_5 BIN_6 BIN_7 BIN_8 BIN_9 BIN_10

It is expected that the other attributes are not set and use printer's default values.

1.2.1.3.5.2 Basic Film Box SOP Class

The ZIOBASE provides the following support for the Basic Film Box attributes sent by the N-CREATE, N-ACTION and N-DELETE service's SCU.

Table 10 : Basic Film Box SOP Class Request Attributes

Attribute	Tag	Valid Range
Image Display Format	(2010,0010)	STANDARD\C,R C=[1..8] R=[1..8]
Film Orientation	(2010,0040)	PORTRAIT LANDSCAPE
Min Density	(2010,0120)	
Max Density	(2010,0130)	
Configuration Information	(2010,0150)	

It is expected that the other attributes are not set and use printer's default values.

1.2.1.3.5.3 Basic Grayscale Image Box SOP Class

The ZIOBASE provides the following support for the Basic Grayscale Image



Box attributes sent by the N-SET service's SCU.

Table 11 : Basic Grayscale Image Box SOP Class Request Attributes

Attribute	Tag	Valid Range
Image Box Position	(2020,0010)	1 - 25
Basic Grayscale Image Sequence	(2020,0110)	
> Samples per Pixel	(0028,0002)	1
> Photometric Interpretation	(0028,0004)	MONOCHROME1 MONOCHROME2
> Rows	(0028,0010)	
> Columns	(0028,0011)	
> Pixel Aspect Ratio	(0028,0034)	1 : 1
> Bits Allocated	(0028,0100)	8 or 16
> Bits Stored	(0028,0101)	8 - 16
> High Bit	(0028,0102)	7 - 15
> Pixel Representation	(0028,0103)	0
> Pixel Data	(7FE0,0010)	
Min Density	(2010,0120)	
Max Density	(2010,0130)	
Configuration Information	(2010,0150)	
Requested Image Size	(2020,0030)	

It is expected that the other attributes are not set and use printer's default values.

1.2.1.3.5.4 Basic Color Image Box SOP Class

The ZIOBASE provides the following support for the Basic Color Image Box attributes sent by the N-SET service's SCU.

Table 12 : Basic Color Image Box SOP Class Request Attributes

Attribute	Tag	Valid Range
Basic Color Image Sequence	(2020,0111)	
> Image Box Position	(2020,0010)	1 - 25
> Samples per Pixel	(0028,0002)	3
> Photometric Interpretation	(0028,0004)	RGB
> Rows	(0028,0010)	
> Columns	(0028,0011)	
> Pixel Aspect Ratio	(0028,0034)	1 : 1
> Bits Allocated	(0028,0100)	8
> Bits Stored	(0028,0101)	8
> High Bit	(0028,0102)	7
> Pixel Representation	(0028,0103)	0



Attribute	Tag	Valid Range
> Pixel Data	(7FE0,0010)	

1.2.1.3.5.5 Printer SOP Class

The ZIOBASE provides the following support for the Printer attributes sent by the N-GET service's SCU.

Table 13 : Printer SOP Class Request Attributes

Attribute	Tag	Valid Range
Printer Status	(2110,0010)	
Printer Status Info	(2110,0020)	
Printer Name	(2110,0030)	
Manufacturer	(0008,0070)	
Manufacturer's Model Name	(0008,1090)	
Device Serial Number	(0018,1000)	
Software Version(s)	(0018,1020)	
Date of Last Calibration	(0018,1200)	
Time of Last Calibration	(0018,1201)	

1.3 Association Acceptance Policy

ZIOBASE accepts the remote AE's associations when all the following conditions are met:

- AE which has established associations is registered in ZIOBASE
- Associations for verification, storage or query/retrieve service
- ZIOBASE is in the association acceptable state

1.3.1 Real-World Activities

1.3.1.1 Description and Sequencing of Real-World Activities

ZIOBASE awaits the associations for Storage, Verification and Query/Retrieve services. If the association proposed from registered AE, it is accepted.

1.3.1.2 Proposed Presentation Contexts

ZIOBASE is capable of proposing the Presentation Contexts shown in the following table.

Table 14 : Proposed Presentation Contexts

Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Verification SOP Class	1.2.840.10008.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
Computed Radiography Image	1.2.840.10008.5.1.4.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Storage		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Digital X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Digital X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.1.1.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Digital Mammography X-Ray Image Storage – For Presentation	1.2.840.10008.5.1.4.1.1.1.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
		Prediction			
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Digital Mammography X-Ray Image Storage – For Processing	1.2.840.10008.5.1.4.1.1.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
CT Image Storage	1.2.840.10008.5.1.4.1.1.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Enhanced CT Image Storage	1.2.840.10008.5.1.4.1.1.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Ultrasound Multi-frame Image Storage	1.2.840.10008.5.1.4.1.1.3.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
MR Image Storage	1.2.840.10008.5.1.4.1.1.4	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Enhanced MR Image Storage	1.2.840.10008.5.1.4.1.1.4.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Ultrasound Image Storage	1.2.840.10008.5.1.4.1.1.6.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
Secondary Capture Image Storage	1.2.840.10008.5.1.4.1.1.7	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
X-Ray Angiographic Image Storage	1.2.840.10008.5.1.4.1.1.12.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
X-Ray Radiofluoroscopic Image Storage	1.2.840.10008.5.1.4.1.1.12.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Nuclear Medicine Image Storage	1.2.840.10008.5.1.4.1.1.20	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None



Presentation Context Table					
Abstract Syntax		Transfer Syntax		Role	Ext. Neg.
Name	UID	Name List	UID List		
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Positron Emission Tomography Image Storage	1.2.840.10008.5.1.4.1.1.128	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
		JPEG Baseline	1.2.840.10008.1.2.4.50	SCP	None
		JPEG Extended	1.2.840.10008.1.2.4.51	SCP	None
		JPEG Lossless, Non-Hierarchical, First-Order Prediction	1.2.840.10008.1.2.4.70	SCP	None
		RLE Lossless	1.2.840.10008.1.2.5	SCP	None
Comprehensive SR Storage	1.2.840.10008.5.1.4.1.1.88.3 3	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
		Explicit VR Little Endian	1.2.840.10008.1.2.1	SCP	None
		Explicit VR Big Endian	1.2.840.10008.1.2.2	SCP	None
Study Root Query/Retrieve Information Model – FIND	1.2.840.10008.5.1.4.1.2.2.1	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None
Study Root Query/Retrieve Information Model – MOVE	1.2.840.10008.5.1.4.1.2.2.2	Implicit VR Little Endian	1.2.840.10008.1.2	SCP	None

1.3.1.3 SOP Specific Conformance

1.3.1.3.1 SOP Specific Conformance for Verification SOP Class

ZIOBASE provides Standard Conformance as SCP of Verification SOP Class.

1.3.1.3.2 SOP Specific Conformance for Storage SOP Class

ZIOBASE can accept multiple C-STORE requests as Storage Service provider with a single association. ZIOBASE is Level 2 (Full) conformant as a Storage Service provider. No attributes are overridden and modified. If C-STORE is successful, ZIOBASE stores the DICOM image data to hard disk. The data stored in a hard disk is enabled to access through the user of the OS service or ZIOBASE.

ZIOBASE returns C-STORE response statuses shown in the following table.



Table 15 : Storage C-STORE Response Status

Error Code	Service Status	Reason
0x0000	Success	Success
0xA700	Refused	Out of Resources
0xA900	Error	Data Set does not match SOP Class
0xB000	Warning	Coercion of Data Elements
0xB006	Warning	Elements Discarded
0xB007	Warning	Data Set does not match SOP Class
0xCxxx	Error	Any other error

1.3.1.3.3 SOP Specific Conformance for Query/Retrieve Information Model – FIND

ZIOBASE does not support relational queries. No extended negotiation is performed. ZIOBASE provides Standard Conformance as the Query SOP Class SCP. ZIOBASE returns C-FIND response statuses shown in the following tables.

Table 16 : Study Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type	Available for Query key
Study Date	(0008,0020)	R	Yes
Study Time	(0008,0030)	R	No
Accession Number	(0008,0050)	R	Yes
Patient's Name	(0010,0010)	R	Yes
Patient ID	(0010,0020)	R	Yes
Study ID	(0020,0010)	R	Yes
Study Instance UID	(0020,000d)	U	Yes
Number of Study Related Series	(0020,1206)	O	No
Number of Study Related Instances	(0020,1208)	O	No
Modalities in Study	(0008,0061)	O	Yes
Referring Physician's Name	(0008,0090)	O	Yes
Study Description	(0008,1030)	O	Yes

Table 17 : Series Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type	Available for Query key
Modality	(0008,0060)	R	Yes
Series Number	(0020,0011)	R	Yes
Study Instance UID	(0020,000d)	U (*)	Yes
Series Instance UID	(0020,000e)	U	Yes
Series Description	(0008,103e)	O	Yes
Body Part Examined	(0018,0015)	O	Yes
Protocol Name	(0018,1030)	O	Yes
Study ID	(0020,0010)	O	Yes
Number of Series Related Instances	(0020,1209)	O	No

(*): Higher level unique key

Table 18 : Composite Object Instance Level Request Attributes for Study Root Query/Retrieve Information Model

Attribute Name	Tag	Type	Available for Query key
Instance Number	(0020,0013)	R	Yes



Attribute Name	Tag	Type	Available for Query key
SOP Instance UID	(0008,0018)	U	Yes
Study Instance UID	(0020,000d)	U(*)	Yes
Series Instance UID	(0020,000e)	U(*)	Yes

(*): Higher level unique key

1.3.1.3.4 SOP Specific Conformance for Query/Retrieve Information Model – MOVE

ZIOBASE does not support relational queries. No extended negotiation is performed. ZIOBASE provides Standard Conformance as SCP of retrieve SOP Class.

1.3.1.4 Proposed Presentation Contexts

ZIOBASE is capable of proposing multiple Presentation Contexts. The specification is shown as follows:

(Abstract Syntax + (Transfer Syntax * Number of Syntax)) * Number of Syntax

In this case, number of syntax is equal to or greater than 1. This capability is configurable.

1.3.1.5 Transfer Syntax Selection Policy

ZIOBASE supports various Transfer Syntaxes. The selection is decided as follows:

1. The priority list of Transfer Syntax is configured.
2. Propose the Transfer Syntax in order.
3. First conformed Transfer Syntax is selected.

2 Network Profile

2.1 Supported Protocol Stacks (PS 3.8, PS 3.9)

ZIOBASE provides DICOM V3.0 TCP/IP Network Protocol Stacks support in which stacks are defined in DICOM Standards PS 3.8.

2.2 OSI Stack

OSI Stack is not supported.

2.3 TCP/IP Stack

ZIOBASE inherits TCP/IP Stack from runtime environment OS.

2.4 API

APIs are not released.



2.5 Physical Device Support

ZIOBASE inherits Physical Device Support from runtime environment OS.

2.6 Point-to-point stack

Point-to-point stack is not supported.

3 STANDARD EXTENDED/SPECIALIZED/PRIVATE SOP CLASSES

ZIOBASE does not claim conformance to any Extended, Specialized or Private SOP Classes.

4 Configuration

4.1 AE Title/Presentation Address Mapping

AE Title/Presentation Address Mapping is set while in installation by the support engineer.

4.2 Parameters

The following parameters are available. These are configured during installation or maintenance by the support engineer.

- AE Title
- IP Address
- Port Number
- Max PDU size
- Priority of the acceptable Transfer Syntax

5 SUPPORT OF CHARACTER SETS

ZIOBASE does not support extended character sets for Query service. Otherwise, the following character sets are supported:

- ISO-8859-1 (ISO-IR 100 Latin-1)
- JIS X 0201 (ISO-IR 13 Japanese katakana and ISO-IR 14 Japanese romaji)
- JIS X 0208 (ISO-IR 87 Japanese kanji, hiragana and katakana)
- JIS X 0212 (ISO-IR 159 Supplementary Japanese kanji)